


# VA test checker FREE CHLORINE



# VA test

www.vagnerpool.com

Dear Customer,

Thank you for choosing Vagner Pool Product.

Please read this instruction manual carefully before using the instrument.

If you need additional technical information, do not hesitate to e-mail us at [info@vagnerpool.com](mailto:info@vagnerpool.com).

### Preliminary examination:

Please examine this product carefully. Make sure that the instrument is not damaged. If any damage occurred during shipment, please notify your Dealer.


Each 690H701 meter is supplied complete with:

- Two Sample Cuvettes and Caps
- Six powder reagents for Free Chlorine
- 1 x 1.5V AAA Battery
- Instruction Manual

### Technical specifications:

Range	0,00 to 2,50 mg/L
Resolution	0,01 ppm
Accuracy	±0,03 mg/L nebo ±3% na odečet při 25°C
Typical EMC Dev.	±0,01 mg/L
Light Source	LED dioda s vlnovou délkou 525 nm
Light Detector	Křemíkový fotodetektor
Method	Adaptation of USEPA method 330.5. The reaction between free chlorine and the DPD reagent causes a pink tint in the sample.
Environment	0 to 50°C (32 to 122°F) max. 95% RH non-condensing
Battery Type	1 x 1.5V AAA
Auto-Shut off	After 2 minutes of non-use and 10 seconds after reading.
Dimensions	81.5 x 61 x 37.5 mm (3.2 x 2.4 x 1.5")
Weight	64 g (2.25 oz.).







### Functional description:




### Errors and warnings:

	<b>Light High:</b> There is too much light to perform a measurement. Please check the preparation of the zero cuvette.
	<b>Light Low:</b> There is not enough light to perform a measurement. Please check the preparation of the zero cuvette.
	<b>Inverted cuvettes:</b> The sample and the zero cuvette are inverted.
	<b>Under range:</b> A blinking "0,00" indicates that the sample absorbs less light than the zero reference. Check the procedure and make sure you use the same cuvettes for reference (zero) and measurement.
	<b>Over Range:</b> A flashing value of the maximum concentration indicates an over range condition. The concentration of the sample is beyond the programmed range; dilute the sample and re-run the test.
	<b>Battery low:</b> The battery must be replaced soon.
	<b>Dead battery:</b> This indicates that the battery is dead and must be replaced. Once this indication is displayed, normal operation of the instrument will be interrupted. Change the battery and restart the meter.

### Measurement procedure:

- Turn the meter on by pressing the button. After all the segments are displayed, "C.1", "Add" appears with "Press" blinking, the meter is ready. 
- Fill the cuvette with 10 mL of unreacted sample and replace the cap. Place the cuvette into the meter and close the meter's cap. 
- Press the button. When the display shows "Add", "C.2" with "Press" blinking the meter is zeroed. 
- Remove the cuvette, open it and add the content of one packet of HI 701-25 reagent. Replace the cap and shake gently for 20 seconds. Replace the cuvette into the meter. 
- Wait for 1 minute and then press the button or press and hold the button until the timer is displayed on the LCD. 
- The instrument directly displays the concentration of free chlorine in ppm. The meter automatically turns off after 10 seconds. 




### Tips for an accurate measurement

- It is important that the sample does not contain any debris.
- Whenever the cuvette is placed into the measurement cell, it must be dry outside, and completely free of fingerprints, oil or dirt. Wipe it thoroughly with HI 731318 or a lint-free cloth prior to insertion.
- Shaking the cuvette can generate bubbles, causing higher readings. To obtain accurate measurements, remove bubbles by swirling or by gently tapping the cuvette.
- Do not let the reacted sample stand for too long after reagent is added, or accuracy will be lost.
- After the reading, it is important to immediately discard the sample, otherwise the glass might become permanently stained.

### Battery management

To save the battery, the instrument shuts down after 2 min. of non-use and 10 sec. after read. One fresh battery lasts for a minimum of 5000 measurements, depending on the light level. When the battery capacity is under 10 % "bAt" appears on the LCD at start-up. If the battery is empty and accurate measurements can't be taken any more, the instrument shows "bAd" then "bAt" each for 1 second and turns off. To restart the instrument, the battery must be replaced with a fresh one. To replace the instrument's battery, follow the steps:

- Turn the instrument off by holding the button until the meter shuts off.
- Turn the instrument upside down and remove the battery cover with a screwdriver.



### Recycling Used Electronic Equipment

Please note that this device or batteries must be disposed of as electronic waste and can be included in used electronic equipment recycling programs.



### Accessories:

Reagents sets  
69HI 9370101 Reagents for 100 Free Chlorine tests

#### Other accessories

- HI 740028 1.5V AAA batteries (4 pcs)
- HI 731318 Tissue for wiping cuvettes (4 pcs)
- HI 731321 Glass cuvettes (4 pcs)
- HI 731225 Caps for cuvettes (4 pcs)
- HI 93703-50 Cuvettes cleaning solution (230 mL)

### Recommendations for Users

Before using these products, make sure that they are entirely suitable for your specific application and for the environment in which they are used. Operation of these instruments may cause unacceptable interferences to other electronic equipments, thus requiring the operator to take all necessary steps to correct interferences.

### Warranty

For checker VA test FREE CHLORINE checker the warranty is 2 years. If you need additional technical information, do not hesitate to contact us.

# VA test

www.vagnerpool.com